

# Middlesex University Research Repository

An open access repository of

Middlesex University research

<http://eprints.mdx.ac.uk>

Liston, Adam D., Bagshaw, Andrew P., Bayford, Richard ORCID logoORCID:  
<https://orcid.org/0000-0001-8863-6385>, Tizzard, Andrew ORCID logoORCID:  
<https://orcid.org/0000-0002-6159-4901>, Tidswell, A.T., Dehghani, Hamid and Holder, David S.  
(2003) Effects of modelling layers and realistic geometry in reconstruction algorithms for EIT of  
brain function. Proceedings of the 4th Conference on Biomedical Applications of Electrical  
Impedance Tomography: Abstracts.. In: 4th Conference on Biomedical Applications of  
Electrical Impedance Tomography, 23-25 Apr 2003, Manchester, UK. . [Conference or  
Workshop Item] (doi:10.5281/zenodo.17924)

First submitted uncorrected version (with author's formatting)

This version is available at: <https://eprints.mdx.ac.uk/2921/>

## Copyright:

Middlesex University Research Repository makes the University's research available electronically.

Copyright and moral rights to this work are retained by the author and/or other copyright owners unless otherwise stated. The work is supplied on the understanding that any use for commercial gain is strictly forbidden. A copy may be downloaded for personal, non-commercial, research or study without prior permission and without charge.

Works, including theses and research projects, may not be reproduced in any format or medium, or extensive quotations taken from them, or their content changed in any way, without first obtaining permission in writing from the copyright holder(s). They may not be sold or exploited commercially in any format or medium without the prior written permission of the copyright holder(s).

Full bibliographic details must be given when referring to, or quoting from full items including the author's name, the title of the work, publication details where relevant (place, publisher, date), pagination, and for theses or dissertations the awarding institution, the degree type awarded, and the date of the award.

If you believe that any material held in the repository infringes copyright law, please contact the Repository Team at Middlesex University via the following email address:

[eprints@mdx.ac.uk](mailto:eprints@mdx.ac.uk)

The item will be removed from the repository while any claim is being investigated.

See also repository copyright: re-use policy: <http://eprints.mdx.ac.uk/policies.html#copy>

# Effects of modelling layers and realistic geometry in reconstruction algorithms for EIT of brain function

**Adam D. Liston<sup>†</sup>, Andrew P. Bagshaw<sup>‡</sup>, Richard H. Bayford<sup>†</sup>, Andrew Tizzard<sup>†</sup>, A. Thomas Tidswell<sup>§</sup>, Hamid Dehghani<sup>#</sup> and David S. Holder<sup>§</sup>**

<sup>†</sup>School of Health and Social Sciences, Middlesex University, London.

<sup>‡</sup>Montreal Neurological Insititute, Montreal, Canada.

<sup>§</sup>Department of Clinical Neurophysiology, University College Hospital, London, UK

<sup>#</sup>Thayer School of Engineering, Dartmouth College, Hanover, USA

Electrical impedance tomography of brain function has the potential to provide a rapid portable bedside neuroimaging device. Recently, our group published the first ever EIT images of evoked activity recorded with scalp electrodes. While the raw data showed encouraging reproducible changes of a few per cent, the images were noisy. The poor image quality was due in part to the use of a simplified reconstruction algorithm which modelled the head as a homogeneous sphere. The purpose of this work has been to develop new algorithms in which the model incorporates extracerebral layers and realistic geometry, and to assess their effect on image quality.

We considered algorithms incorporating four different forward models of the head: two analytical models i) a homogeneous sphere and ii) concentric spheres in which the brain, skull and scalp were modelled analytically and two numerical (FEM) models iii) a homogeneous, head-shaped volume and iv) a head-shaped volume with internal compartments of contrasting resistivity.

The two analytical sphere algorithms were tested on a) a computer-simulation using the four, concentric spheres model; b) a spherical, electrolyte-filled tank containing a concentric Plaster of Paris shell to mimic the skull. Mean localisation errors for 38 simulated perturbations were  $5.82.1\text{mm}$  and  $15.96.3\text{mm}$  respectively for four-shell and homogeneous reconstruction. For movement of Perspex along the same axes in the tank, mean localisation errors were  $13.7 \pm 5.8\text{mm}$  and  $20.7 \pm 10.1\text{mm}$ . Incorporating shells in the algorithm resulted in peak impedance changes localised more accurately and less centrally than when shells were not included.

All four algorithms were tested on 8 human subjects during between 6 and 12 epochs of exposure to a visual stimulus. Images of visual evoked responses showed no more consistency across subjects with the new algorithms though some individual examples were improved.